

22.26 *Replacement:*

The output circuit shall be supplied through a **safety isolating transformer** and shall not be connected to **accessible metal parts** or an earthing terminal. The insulation between parts operating at **safety extra-low voltage** and **live parts** shall comply with the requirements for **double insulation** or **reinforced insulation**.

*Compliance is checked by inspection and by the tests specified for **double insulation** or **reinforced insulation**.*